



SERTIKA
CERTIFICATION CENTER



LIETUVOS
NACIONALINIS
AKREDITACIJOS
BIURAS

PRODUKTŲ SERTIFIKAVIMAS
EN 45011 Nr. LA.03.002

CERTIFICATE

Certificate Registration No. LS.08.02.2105

24 January, 2011

Aerosol Halogenerator

HaloPrima-01 (ASG P-01)

Made by "AEROMED-GENESUNG" GmbH
Plaßstraße 66B 33611 Bielefeld, Germany

Conforms to the requirements of the standards:
EN 60601-1:2002, IEC 61010:2001
EN 60601-1-2:2007 (EN 61000-4-4, EN 61000-4-5,
EN 61000-3-2, EN 55011).

Certificate is issued on the base of:

Test Reports: No. 05-10S, issued by Testing Laboratory of Certification Center SERTIKA, No. 29.1 PB-126 issued by Equipment and Devices Electromagnetic Compatibility Control Section of Communications Regulatory Authority (LITHUANIA).

Certificate gives the right to mark the certified product with the conformity mark of SERTIKA.

This certificate is valid up to 24 January, 2012.

Head of Product
Certification Service



Paulius Dambrauskas